SCT Ambulance Equipment List July 1, 2022 This equipment list is not intended to provide a complete description of every item required by COMAR 30.09

Service:			_		Fleet ID:
Locati	on:		-		Inspector:
Line	No. of	Description	Pass	Fail	Notes
#:	Items	All ALS ambulance licensure requirements must be met including any approved OSP's.			
		Infusion Pumps	1	ı	
1	4	infusion pumps			
2	6	infusion pump tubing sets			
3	4	extension sets			
4	4	stopcocks			
5	2	blood infusion sets			
6	2	intravenous filters			
Ventilator					
7	1	ventilator minimally capable of the following settings:			
8		oxygen concentration between 20% - 100%			
9		peak pressures between 0 and 99			
10		selectable intermittent mandatory ventilation (IMV)			
11		positive end expiratory pressure between 2 and 15			
12		adjustable ventilatory rated between 0 and 60 per minute			
13		adjustable tidal volumes:			
14		adults: 300ml - 1,000ml			
15		pediatric: 15mg - 200ml			
16		adjustable high and low pressure alarms			
17	2	ventilator circuits			
Cardiac Monitor					
18	1	cardiac monitor capable of monitoring:			
19		invasive blood pressure			
20		ETCO2 with waveform display & oxygen saturation			
Adult Transports					
21	1	surgical crichthryotomy kit			
22	1	needle crichthryotomy kit			
23	2	surgical clamps capable of clamping a minimum of a 40F chest tube or Foley			
		Pediatric Transports	Г	Γ	
24		500mL of D5 Normal Saline or Normal Saline maintenance fluid			
25	6	1cc syringes			
26	1	surgical crichthryotomy kit			
27	1	needle chrichthryotomy kit Temperature Monitoring			
28	1	device to monitor induced hypothermia	Ī		
29	1	electric thermometer			
30	1	device to monitor core temperature (rectal or esophageal)			
Medications					
31		all additional medications approved for use by the specialty care service Medical Director to meet the needs of the patients being transported			
32		a current list of all approved medication must be available on each ambulance when used as SCT			
33		sufficient quantities of medications shall be present to care for on adult for the longer of one hour			