

ALS – Ambulance

Company: _____ Fleet ID: _____

VIN: _____ Inspector: _____

Inspection Date: _____ Needs Decal: Yes/No (Quantity: _____)

Deficiencies	Corrected

Line #	Quantity	Description	Pass	Fail	Notes
1		BLS Ambulance Seal of Excellence Requirements met.			
ALS Equipment					
2	1 ea.	Cardiac monitor/defibrillator with the following capabilities: <ul style="list-style-type: none"> Quick look capability (adult and pediatric) Waveform capnography Ability to transfer data into eMEDS® record-transmission If unit is BLS staffed regularly, an AED with pediatric capability is recommended. 			
3	2 ea.	Multi-function pads (adult) <i>(check exp. dates)</i>			
4	2 ea.	Multi-function pads (pediatric) <i>(check exp. dates)</i>			
5	1 set	Monitoring cables			
6	30	Monitoring electrodes			
7	1 ea.	Spare monitor/defibrillator batteries and/or on-board charging system			
8	1 roll	Spare EKG paper			
9	1 ea.	Pulse oximeter			
10	1 ea.	ICD donut magnet			
11	1 ea.	CPAP device (circuit connecting directly to O2 is acceptable)			
12	2 ea.	CPAP masks/circuits and in-line nebulizers			
13	1 ea.	Ventilator (if participating in optional protocol)			
14	1 ea.	Portable ultrasound device (if participating in pilot protocol)			
15	-	Gastric tubes (8 Fr and various sizes 10 – 16 Fr adult) Feeding tubes are acceptable. Suction catheters (usually #8) are acceptable if thumb hole can be occluded. Minimum of 3 sizes recommended. <i>(check exp. dates)</i>			
16	1 ea.	Tapered tip lavage syringe – 30 mL (minimum)			
17	2 ea.	14g x 3.25” needle for NDT			
18	1 ea.	Pediatric reference guide			
Medication and Delivery Devices <i>(check for exp. dates)</i>					
Packing of medications or IV solutions may vary, but quantities must be met					
20	30 mg	Adenosine			
21	20 mg	Albuterol sulfate			
22	600 mg	Amiodarone			
23	325mg	Aspirin (chewable)			

ALS – Ambulance

Medication and Delivery Devices (continued)

24	1 mg	Atropine sulfate			
25	8 mg	Atropine sulfate, multi-dose vial			
26	2 g	Calcium chloride			
27	3 g	Calcium gluconate (OPTIONAL)			
28	10 mg	Dexamethasone			
29	500 mL	Dextrose 10% solution (Must carry D50 if D10 is unavailable)			
30	50 g	Dextrose 50% (OPTIONAL)			
31	50 mg	Diltiazem (If using lypholized, at least 1 appropriate mini-bag required)			
32	50 mg	Diphenhydramine			
33	400 mg	Dopamine (OPTIONAL - if approved by OMD)			
34	5mg	Droperidol			
35	6 mg	Epinephrine 0.1mg/mL (prev 1:10,000)			
36	3 mg	Epinephrine 1 mg/mL (prev 1:1,000)			
37	40 mg	Etomidate (if participating in RSI pilot)			
38	400mcg	Fentanyl: must be under lock or controlled access system.			
39	3 ea.	Glucagon – 1mg each			
40	10 mg	Haloperidol (OPTIONAL - if approved by OMD for Droperidol shortage)			
41	-	Hydroxycobalamin (OPTIONAL)			
42	500 mcg	Ipratropium bromide			
43	800 mg	Ketamine: must be under lock or controlled access system.			
44	30mg	Ketorolac (OPTIONAL)			
43	100 mg	Lidocaine 2%			
45	4 mL	Lidocaine 4% (OPTIONAL)			
46	4 g	Magnesium sulfate			
47	20 mg	Midazolam: must be under lock or controlled access system.			
48	40 mg	Morphine sulfate (OPTIONAL): must be under lock or controlled access system.			
49	6 mg	Naloxone			
50	1 ea.	Nitroglycerin spray/tablet bottle (or 3 individually packaged 0.4mg tablets)			
51	1 g	Nitro paste and applicator			
52	24 mg	Ondansetron (vial or ODT)			
53	150 mEq	Sodium bicarbonate			
54	200 mg	Succinylcholine (if participating in RSI pilot)			
55	1g	Tranexamic Acid (TXA)			
56	10 mg	Vecuronium (if participating in RSI pilot)			
57	10 mg	Verapamil (OPTIONAL - if approved by OMD for Diltiazem shortage)			
58	2 ea.	Nebulizers			
59	1 ea.	Controlled access system: DEA-controlled substances must be under lock or controlled access system.			
60	4 ea.	1 mL syringes			
61	2 ea.	3-5 mL syringes			
62	2 ea.	18g or 19g blunt needles			
63	2 ea.	21g needles (1.5” for IM injection)			
64	2 ea.	10 mL syringes			
65	2 ea.	Intranasal medication delivery devices – (check exp. dates)			

ALS – Ambulance

Intravenous Equipment and Supplies (check for exp. dates)

66	4 ea.	IV catheters (14g, 16g, 18g, 20g, 22g, 24g)			
67	-	IO needles (2-15g and 2-18g if manual) or (1-15mm, 1-25mm, 1-45mm and 1-18g jamshidi if mechanical)			
68	3 sets	IV admin. sets (2 capable of 10 – 15 drops/min. and 1 capable of 60 drops/min.)			
69	2ea.	Normal saline, LR, or D5W: 100cc bag			
70	-	IV bags Lactated Ringer’s (Must have 4,000 cc / at least 1 1,000cc bag)			
71	-	Site preparation materials			
72	1 ea.	Adult IV arm board (Short board splint can substitute)			
73	1 ea.	Pediatric IV arm board (max width 2”)			
74	2 ea.	Saline flush			
75	2 ea.	Saline lock			
76	2 sets	Blood draw supplies (OPTIONAL Jurisdictional Requirement) <ul style="list-style-type: none"> • 3 ea. - blood tubes (type varies by jurisdiction) (check exp. date) • 3 ea. - blood tubes with anticoagulant (type varies by jurisdiction) (check exp. date) • Vacutainer 			
77	1 ea.	Portable sharps container			
78	2 ea.	3-way stop cocks with extension tubing			
79	2 ea.	Non-coring right angle needles (i.e., Huber Needles)			

Intubation Kit (check for exp. dates)

80	1 set	Miller blades (0, 1, 2, 3, 4)			
81	1 set	Macintosh blades (1, 2, 3, 4)			
82	1 ea.	Large laryngoscope handle with spare batteries			
83	1 ea.	Small laryngoscope handle with spare batteries (OPTIONAL)			
84	1 ea.	Spare laryngoscope bulbs (OPTIONAL)			
85	2 ea.	ET tubes cuffed (6, 7, 8)			
86	2 ea.	ET tubes cuffed or uncuffed (2.5, 3, 3.5, 4, 4.5, 5, 5.5) (Cuffed tubes recommended)			
87	1 ea.	Flexible tracheal tube guide (i.e., Gum Elastic Bougie) (OPTIONAL) (Required July 1, 2024)			
88	2 ea.	Adult stylette			
89	2 ea.	Pediatric/infant stylette			
90	1 roll	1” medical tape			
91	2 ea.	10 mL syringes			
92	1 ea.	Magill forceps (large)			
93	1 ea.	Magill forceps (small)			
94	1 ea.	Water soluble lubricant (1 tube or 3 packets)			
95	1 ea.	ETCO2 (electronic) pediatric and adult			
96	1 set	Extraglottic airway device (One of the following: iGel, LMA, Air-Q, or King-LT; all sizes per protocol)			
97	1 ea.	ET tube holders			
98	1 ea.	Suction device for meconium			
99	1 ea.	Beck Airway Airflow Monitor (BAAM)			
100	1 ea.	Video laryngoscope with recording device (OPTIONAL)			

