

BLS First Responder Unit

Company: _____ Fleet ID: _____

VIN: _____ Inspector: _____

Inspection Date: _____ Needs Decal: Yes/No (Quantity: _____)

Deficiencies	Corrected

Line #	Quantity	Description	Pass	Fail	Notes
General Supplies					
1	1 ea.	Blanket			
2	1 ea.	Obstetrical (OB) Kit (commercially packaged)			
3	2 ea.	Clean linen sheets or Mylar blanket (for burns)			
4	1 ea.	Maryland Triage Kit: Enough triage tags and color-coded ribbons to triage 25 patients. Kits must contain paperwork for the treatment and transportation areas.			
5	–	ANSI 207-2006 class II reflective safety vests for each crew member			
6	1 ea.	PHMSA Emergency Response Guidebook (ERG): Current edition in either print or electronic format			
7	1 ea.	Environmental carbon monoxide alarming device (OPTIONAL)(RECOMMENDED)			
Portable First Aid Kit					
8	12 ea.	Sterile gauze pads (min. 4" x 4")			
9	4 ea.	Sterile dressings (min. 5" x 9")			
10	1 ea.	Hemostatic agent/Hemostatic gauze (OPTIONAL): Chitosan or kaolin based and must be FDA approved/cleared.			
11	4 rolls	Self-adhering gauze bandages (various sizes 2" - 6")			
12	4 ea.	Cravats (triangular bandages): Minimum size is 36" x 36".			
13	2 ea.	Cold packs			
14	1 ea.	Bottle normal saline and/or sterile water (500cc): (check exp. date)			
15	4 ea.	Commercially available tourniquet capable of stopping arterial blood flow. (Must be a CoTCCC recommended device.)			
16	2 rolls	1" medical tape (hypoallergenic tape must be available)			
17	1 ea.	Ring cutter			
18	1 ea.	Bandage scissors or rescue shears (at least 5.5")			
19	1 ea.	Penlight (narrow beam flashlight acceptable)			
20	1 ea.	Stethoscope (must be pediatric-capable)			
21	1 ea.	Glucometer kit – must include lancets, test strips, alcohol wipes, and band-aids. (check exp. date on test strips.)			

BLS First Responder Unit					
Portable First Aid Kit (continued)					
22	1 ea.	Adult Regular BP cuff			
23	1 ea.	Adult Large BP Cuff			
24	1 ea.	Child BP Cuff			
25	6 pairs	Non-latex exam gloves (assorted sizes)			
26	1 ea.	Portable sharps container			
27	1 ea.	Pulse oximeter (Adult and Peds sensors recommended)			
28	1 ea.	Clean kit large enough to carry all supplies in "Portable First Aid Kit"			
Medications and Delivery Devices (check expiration dates)					
29	5mg	Albuterol sulfate- may substitute one albuterol hand-held inhaler			
30	325mg	Aspirin (chewable)			
31	2 ea.	Glucose paste tubes			
32	6ea.	Isopropyl Alcohol Wipes			
33	4mg	Naloxone (Narcan)			
34	2 ea.	Intranasal medication delivery device (not necessary if pre-packaged with naloxone)			
35	1 ea.	Epinephrine auto-injector (adult): Draw-up epi may be substituted if jurisdiction is approved for the optional protocol			
36	1 ea.	Epinephrine auto-injector (pediatric): Draw-up epi may be substituted if jurisdiction is approved for the optional protocol			
37	-	DuoDote (OPTIONAL)			
AED					
38	1 ea.	AED – pediatric cable			
39	2 ea.	AED pads (adult) – (check exp. dates)			
40	1 ea.	AED pads (pediatric) – (check exp. date)(Only required if device or pads does not have dose attenuator)			
41	1 ea.	Spare battery (if required)			
42	1 ea.	Razor			
43	1 ea.	Washcloth or towel (OPTIONAL)			
Biohazard Items					
44	1ea.	Surgical masks: Should be provided for each seated position on the unit. Must have one for patient.			
45	1ea.	Gowns (impenetrable to body fluids): Should be provided for each seated position on the unit.			
46	1ea.	Eye/facial shield (may be combined with mask): Should be provided for each seated position on the unit.			
47	1ea.	Particulate respirator N95 or greater for each crew member: N95 needs to be fit tested for a proper fit. Should be provided for each seated position on the unit.			
48	1 ea.	Appropriate disinfectant: Effective against pathogens to include hepatitis B virus (HBV), human immunodeficiency virus (HIV), and M. tuberculosis (TB), and SARS-CoV2. Minimum of wipes are sufficient. (check exp. date)			

BLS First Responder Unit

Portable Oxygen Kit

49	2 ea.	Nasal cannula (adult)			
50	2 ea.	Nasal cannula (pediatric)			
51	2 ea.	Non-rebreather mask (adult)			
52	2 ea.	Non-rebreather mask (pediatric)			
53	1 ea.	Nebulizers - not necessary if using hand-held albuterol inhaler			
54	1 ea.	Adult (1,000 – 1,200 mL) hand-operated, self re-expanding bag resuscitator without a pop-off valve or with a selectable pop-off valve. Peep valve (REQUIRED) and viral filter recommended.			
		An oxygen inlet			
		Reservoir tube			
55	1 ea.	Transparent face mask (adult)			
56	1 ea.	Child (750 mL) hand-operated, self re-expanding, bag resuscitator without a pop-off valve or with a selectable pop-off valve. PEEP valve (REQUIRED) and viral filter recommended.			
		An oxygen inlet			
		Reservoir tube			
57	1 ea.	Infant (450 - 500 mL) hand-operated, self re-expanding, bag resuscitator without a pop-off valve or with a selectable pop-off valve. PEEP valve (REQUIRED) and viral filter recommended.			
		An oxygen inlet			
		Reservoir tube			
58	1 set	Transparent face mask (neonate through small adult; a set is 4 sizes): Masks can be included with BVM			
59	1 set	Oropharyngeal airway (newborn through large adult; a set is 6 sizes)			
60	1 set	Nasopharyngeal airways (18 Fr – 34 Fr., a set is 6 sizes): (check exp. dates)			
61	1 ea.	Water soluble lubricant (1 tube or 3 packets): (check exp. dates)			
62	1 ea.	Medical oxygen cylinder with at least 300 L capacity (required “E,” “D,” or super D size): Portable tank minimum of 500 psi. Portable tank must be in a portable bag that can be carried to patient side.			
		(#1) Year PSI			
		Cylinder properly color-coded (green = steel, unpainted = brushed metal for aluminum or stainless steel)			
		Free of grease, oil, or other flammable organic material			
		Passed hydrostatic testing within the past 5 years for aluminum and steel cylinders: Steel cylinders with a hydrostatic test date followed by a star is acceptable for 10 years.			

BLS First Responder Unit

Portable Oxygen Kit (continued)

		Regulator shall have a pressure gauge to indicate the pressure of oxygen remaining in the cylinder (not gravity-dependent). Can be separate or combined with oxygen pressure gauge.			
		A variable flow valve and a flowmeter capable of delivering at least 15 LPM, with a dial-down rate to a minimum of 2 LPM			
		Accurate within 1 LPM when setting equal to or less than 5 LPM Test reading of __LPM when flowmeter set at 4 LPM (3 – 5 LPM)			
		Accurate within 1.5 LPM when setting between 6 and 10 LPM Test reading of __LPM when flowmeter set at 10 LPM (8.5 – 11.5 LPM)			
		Accurate within 2 LPM when setting equal to or greater than 15 LPM Test reading of __LPM when flowmeter set at 15 LPM (13 – 17 LPM)			
63	1 ea.	Clean kit large enough to carry all supplies in “Portable Oxygen Kit”			

Portable Suction Unit

64	1 ea.	Manual pumps must meet the same testing requirements as a battery-operated suction device, and have the following:			
64a		All of the required manufacturer parts			
64b		Adult soft tip catheter			
64c		Adult hard tip catheter			
64d		Pediatric soft catheter			
65		If using battery-powered suction unit, it must be capable of operating continuously under suction for at least 20 minutes with a rigid suction tip			
65a		Must be able to develop 11.81 inches of water vacuum (300 mmHg) within 4 seconds of clamping Test reading at 4 sec. ____in/Hg			
65b		A free air flow of at least 20 LPM at the end of the suction tube. Test reading ____LPM			
65c	-	Assorted catheters 6 – 18 Fr and rigid suction tips: One must be between 6 and 12 Fr AND one between 12 and 18 Fr (check exp. dates)			

First Responder Unit Vehicle

66		Functional emergency warning lights			
67		Functional emergency audible warning devices (not horn)			
68		Functional head, tail, and signal lights			
69	1 ea.	Fire extinguisher (5 lb. multipurpose dry chemical). Check label or bottom of cylinder for date. Must be mounted or secured. May be mounted in outside compartment.			

BLS First Responder Unit

First Responder Unit Vehicle (continued)

70	3 ea.	Reflective road hazard triangles or 3 small traffic cones (Flares are not an acceptable substitute)			
71	1 ea.	Radio w/ capability to communicate with PSAP/EMRC			